

Online Research @ Cardiff

This is an Open Access document downloaded from ORCA, Cardiff University's institutional repository: <https://orca.cardiff.ac.uk/107094/>

This is the author's version of a work that was submitted to / accepted for publication.

Citation for final published version:

Grassi, Elisa, Fioroni, Federica, Berenato, Salvatore, Patterson, Nick, Ferri, Valentina, Braglia, Luca, Filice, Angelina, Versari, Annibale, Iori, Mauro and Spezi, Emiliano 2018. Effect of image registration on 3D absorbed dose calculations in 177 Lu-DOTATOC Peptide Receptor Radionuclide Therapy. *Physica Medica: European Journal of Medical Physics* 45 , pp. 177-185. 10.1016/j.ejmp.2017.11.021 file

Publishers page: <https://doi.org/10.1016/j.ejmp.2017.11.021>
<<https://doi.org/10.1016/j.ejmp.2017.11.021>>

Please note:

Changes made as a result of publishing processes such as copy-editing, formatting and page numbers may not be reflected in this version. For the definitive version of this publication, please refer to the published source. You are advised to consult the publisher's version if you wish to cite this paper.

This version is being made available in accordance with publisher policies.

See

<http://orca.cf.ac.uk/policies.html> for usage policies. Copyright and moral rights for publications made available in ORCA are retained by the copyright holders.



1 *Effect of image registration on 3D **absorbed dose** calculations in ¹⁷⁷Lu-DOTATOC Peptide Receptor*

2 *Radionuclide Therapy*

3

4 Grassi Elisa¹, Fioroni Federica¹, **Berenato Salvatore**⁴, Patterson Nick³, Ferri Valentina¹, **Braglia Luca**⁵, Filice
5 Angelina², Versari Annibale², Iori Mauro¹, Spezi Emiliano^{3,4}.

6

7 Affiliations:

8 1. Medical Physics Unit, **Azienda Unità Sanitaria Locale –IRCCS di Reggio Emilia**, Reggio Emilia, Italy

9 2. Nuclear Medicine Unit, **Azienda Unità Sanitaria Locale –IRCCS di Reggio Emilia**, Reggio Emilia, Italy

10 3. Department of Medical Physics, Velindre Cancer Centre, Cardiff, Wales, UK

11 4. School of Engineering, Cardiff University, Cardiff, Wales, UK

12 5. Unità Studi Clinici e Statistica, **Azienda Unità Sanitaria Locale –IRCCS di Reggio Emilia**, Reggio Emilia,
13 Italy

14

15

16 For correspondence please contact: Elisa Grassi, MS, Medical Physics Unit, **Azienda Unità Sanitaria Locale –IRCCS**
17 **di Reggio Emilia**, Viale Risorgimento 57, 42123 Reggio Emilia; e-mail: elisa.grassi@ausl.re.it

18

1 **Running Title**

2 Effect of image registration in PRRT 3D **absorbed dose** calculation.

1 Abstract

2 Peptide receptor radionuclide therapy (PRRT) is an effective MRT (molecular radiotherapy) treatment, which consists
3 of multiple administrations of a radiopharmaceutical labelled with ^{177}Lu or ^{90}Y . Through sequential functional
4 imaging a patient specific voxel-based time-integrated-activity curve can be derived and used to calculate the
5 cumulated activity. Multiple scans should be co-registered to allow accurate, patient specific 3D dosimetry. The
6 purpose of this study is to measure the quality and to evaluate the impact of image registration algorithms on 3D
7 absorbed dose calculation.

8 A cohort of 11 patients was extracted from the database of a clinical trial in PRRT. They were administered with a
9 single administration of ^{177}Lu –DOTATOC. All patients underwent 5 SPECT/CT sequential scans at 1h, 4h, 24h, 40h,
10 70h post-injection that were subsequently registered using rigid and deformable algorithms. A similarity index was
11 calculated to measure the quality of rigid and deformable registrations and to compare the two algorithms. 3D
12 absorbed dose calculation was carried out with the Raydose Monte Carlo code.

13 The similarity analysis showed that deformable registrations provide superior results than rigid registrations
14 ($p < 0.001$).

15 Average absorbed dose to the kidneys calculated using rigid image registration was consistently lower than the
16 average absorbed dose calculated using the deformable algorithm (90% of cases), with percentage differences in the
17 range [-19;+4]%. Absorbed dose to lesions were also consistently lower (90% of cases) when calculated with rigid
18 image registration with absorbed dose differences in the range [-67.2,100.7]%. Deformable image registration had a
19 significant role in calculating 3D absorbed dose to organs or lesions with volumes smaller than 100mL.

1 Image based 3D dosimetry for ^{177}Lu -DOTATOC PRRT is significantly affected by the type of algorithm used to
2 register sequential SPECT/CT scans. It is advisable to implement deformable image registration in clinical practice.

3

1 Introduction

2 Peptide receptor radionuclide therapy (PRRT) with radiolabelled somatostatin analogues is a specific type of
3 molecular radiotherapy (MRT) and an effective treatment for patients with neuroendocrine tumours (NETs) [1]. It has
4 been shown that three-dimensional (3D) voxel-based patient specific dosimetry is possible and allows to accurately
5 assess the dose absorbed by the organs at risk (OARs) and by the different areas of the disease, as a measure of the
6 potential therapeutic effect [2, 3, 4, 5].

7 Every methodology adopted for dosimetry is affected by a number of metrological issues and uncertainties as
8 highlighted by D'Arienzo et al [6] and by Gustafsson et al [7]. These include, among others, measurement of
9 administered activity, quantification of activity from imaging within a volume in the patient, construction and
10 integration of the time-activity curve from imaging, and calculation of the absorbed dose. Moreover, patient specific
11 pharmacokinetics and individual variations in anatomy determine unique three-dimensional radiopharmaceutical
12 distributions inside the body as a function of time. For PRRT delivered with ^{177}Lu labelled somatostatin receptors the
13 bio-distribution of the radiopharmaceutical may be assessed by performing multiple SPECT/CT scans. Through
14 sequential SPECT/CT scans a patient specific voxel-based time-integrated activity curve can be derived and used to
15 calculate the cumulated absorbed dose with either dose point kernel algorithms or Monte Carlo simulations or local
16 energy deposition.

17 When dealing with sequential functional imaging, the misalignment of sequential scans is a critical aspect, by which
18 the global accuracy of the dosimetry calculations can be strongly affected. Misregistration errors can derive from
19 changes in patient repositioning, organ deformation, tumour progression/regression between different scans and

1 respiratory motion, as reported also by some critical works related to PET/CT diagnostics [8] and external beam
2 radiotherapy [9, 10]. Whatever the source of these errors, they are likely to lead to a poor estimation of organ absorbed
3 doses at the voxel level, and consequently to a sub-optimal treatment plan and to an erroneous prediction of MRT
4 treatment response. On the other hand, several authors recommend to optimize MRT planning to limit the absorbed
5 dose to OARs, while maximizing the tumour control in individualized MRT treatments [3]. Since only a series of 3D
6 images may provide scientists for the possibility to apply all modelling and fitting methods to each voxel individually,
7 approaching a totally personalized treatment, the registration of images should be as more adequate as possible.

8 For these reasons, and together with the need to implement more accurate 3D patient specific dosimetry calculation in
9 MRT, a strong interest towards issues related to image registration has lately raised [11, 12]. Recently Jackson et al.
10 [2] used sequential rigid and deformable registration to align SPECT/CT images in PRRT dosimetry calculations.

11 They performed a CT to CT registration, due to the intrinsic variation in activity distribution of SPECT images, which
12 makes SPECT to SPECT registration inadequate.

13 Although a full comparison between deformable and rigid registration methods was not carried out in that study, the
14 authors concluded that registration accuracy and limited spatial resolution of the SPECT camera may preclude
15 absorbed dose calculations to very small volumes and may have a low impact on estimated absorbed doses of large
16 organs, especially if these organs exhibit homogenous uptake areas.

17 Ao et al [13], on the other hand, didn't present a complete study. They showed that deformable image registration can
18 affect the accuracy of 3D dosimetry with $^{111}\text{In-Zevalin}$, compared to rigid registration, but they based their study on a
19 modelled phantom population. Therefore, important limitations were raised even by the authors: the respiratory

1 motion during SPECT/CT acquisitions was not modelled in the simulation; the study was basically on phantom, so
2 any clinical reference on the realistic degree of deformation among different imaging time points was not considered;
3 they unrealistically simulated uniform activity distributions in organs.

4 Furthermore, their study used only a single patient case (4 SPECT scanned at different time points and 1 CT scan
5 acquired at 24h p.i. for attenuation correction) with the aim to prove that the clinical results were consistent with the
6 simulations.

7 They focused on the impact of organ-by-organ deformable registration on quantitative SPECT images and
8 investigated only the improvement in organ absorbed dose assessment.

9 In this study, instead, we systematically investigate the impact of using image registration within the context of an
10 image based dosimetry trial in PRRT delivered with the administration of a therapeutical activity of ^{177}Lu -DOTATOC
11 in patients suffering from metastasized NET with intense somatostatin receptor expression. Our clinical study
12 included the acquisition of sequential quantitative SPECT/CT scans and the segmentation of both OARs and target
13 lesions in the abdomen (mainly liver) and in the lung. With this work, we aim at demonstrating that it is feasible to
14 implement a deformable image registration workflow in clinical MRT dosimetry and we hypothesize that the use of
15 deformable image registration instead of rigid image registration (as defined by Brock et al [14]) delivers a
16 significantly different representation of the 3D distribution in MRT for OARs and lesions different in size, shape and
17 uptake level.

18 19 Material and Methods

1 Study protocol

2 A group of patients suffering from NET cancer was recruited for PRRT as part of a clinical trial (EUDRACT 2013-
3 002605-65) at AUSL-IRCCS, Reggio Emilia, Italy. The study was conducted in accordance with the provisions of the
4 Declaration of Helsinki and the ICH-GCP Guidelines. All patients entering the trial were asked to give written
5 informed consent for research purposes. Patients were administered **with a therapeutical dosage** of the
6 radiopharmaceutical ^{177}Lu -DOTATOC to perform the internal dosimetry study, according to the EANM guidelines
7 **[15]. The dosimetry study was performed only once after the first injection, though later in time patients received**
8 **multiple fractions of radiopharmaceutical (^{177}Lu -DOTATOC or ^{90}Lu -DOTATOC). The first injected activity was**
9 **chosen based on clinical conditions of patients, since the dosimetry study outcome had not been performed yet. The**
10 **following injected activities were chosen based on more precise patient specific characteristics (general clinical state;**
11 **presence of risk factors such as hypertension, diabetes, previous renal failure; one kidney resection). No injected**
12 **activity was fixed or chosen based only on weight or body surface area, or similar approach, because the**
13 **personalization of PRRT was the aim of the trial.** The adopted administration protocol was described elsewhere [16].

14 A cohort of 11 clinical cases (in the following referred to as the cohort) was extracted from the clinical trial database
15 and considered in this investigation. The choice of the clinical cases was led by the size of the lesions (typical volumes
16 in the range [4mL, 150mL]) and by the anatomical area in which the lesions were located (e.g. different segments of
17 the liver).

18 The mean administered activity for the cohort was 5.3 ± 0.9 GBq. Differences in administered activity were due to
19 differences in patient's weight, height and other clinical parameters. After injection, 5 sequential SPECT/CT

1 abdominal scans (5 SPECT and 5 CT, i.e. one SPECT/CT for each time point; each CT image was used for
2 attenuation correction of the relative SPECT image and for anatomical reference) were taken with the patient in supine
3 position at 1, 4, 24, 44 and 72 hours post injection (p.i.). The standard imaging protocol required that patients kept
4 their arms raised and placed above the head. However, no specific patient positioning protocol was adopted. Patients
5 with special needs were provided with comfort devices such as knee fix or comfort cushions and the same devices
6 were adopted in all sequential images.

7 The abdomen SPECT images were acquired with a dual head gamma camera (SymbiaT2, Siemens Medical, Germany,
8 3/8" NaI(Tl)-detector) and the following settings: two medium energy (ME) collimators; matrix = 128 x 128; zoom =
9 1; views = 32 x 2; time/view = 30 s; step and shoot mode; degree of rotation = 180°; non-circular orbit; detector
10 configuration = 180°. The energy windows (EW) of ¹⁷⁷Lu photo-peaks **used for imaging [17]** were set at 113 keV ±
11 7.5% and 208.4 keV ± 7.5%. For the lower EW, the triple energy window (TEW) scatter correction was used (lower
12 scatter window 87.58 – 104.53 keV, weight = 0.5; the upper scatter window 121.47 - 130.51 keV, weight= 0.9375).
13 For the higher EW, the double energy window (DEW) scatter correction was used (lower scatter window 171.60 -
14 192.40 keV, weight = 0.75). The helical CT parameters were 130kV of voltage with care dose tool activated to
15 optimise the anodic current for patient **dose** saving (maximum anodic current set at 90mAs for the first high quality
16 CT at 1h p.i. and at 40mAs for CT acquired at 4h, 24h, 40h, and 70h p.i.), slice thickness 5mm.

17 The SPECT projections were reconstructed using an iterative algorithm with compensations for attenuation, scatter,
18 and full collimator-detector response as implemented in the E-Soft workstation v32B (Syngo, Siemens Medical
19 Solution, Germany) with the Flash 3D iterative algorithm (10 iterations; 8 subsets; Gaussian filter cut-off = 4.8 mm;

1 4.8 mm cubic voxel). The SPECT/CT acquisition and reconstruction protocols were previously validated by Grassi et
2 al [18, 19].

3 Clinical practice of dosimetry in MRT requires the implementation of a specific calibration protocol allowing for the
4 SPECT image to be expressed in units of Bq/mL [20], since this modality is not at present an intrinsically quantitative
5 imaging technique. In our study a calibration factor (CF) in unit of Bq/counts was derived in reference conditions to
6 convert the SPECT count data in absolute activity per voxel. We used a standard cylindrical plastic phantom of 5640
7 mL volume (Data Spectrum Corporation, Durham, USA) filled with a homogenous solution of ^{177}Lu (0.25 MBq/mL).
8 The calibration protocol was described in detail elsewhere [18]. In brief, the reference phantom was scanned 5 times
9 after preparation until the activity concentration reached a value of approximately 0.0078 MBq/mL (~ 5 half-lives of
10 ^{177}Lu). As expected from the results about dead time obtained by Grassi et al [19], the CF was not affected by the
11 count rate in the range of activity considered. The CF values calculated at different time points showed no significant
12 deviation from a linear trend, and had a mean value of 28.5 Bq/counts and a standard error of 4%. This value was used
13 to convert the clinical SPECT images in absolute activity distributions. **Our scanner was characterized for Partial**
14 **Volume Effect (PVE) [18,19] but contrast recovery was not considered here.**

15 For each patient, both right and left kidneys were manually segmented on the first CT scan by a nuclear medicine
16 physician together with a nuclear medicine physicist using the VoxelMed software [18]. **This** CT scan was acquired
17 with no image contrast agent and with a high image quality protocol. In addition, target lesions were manually
18 outlined with the same software on the same CT scan when possible or alternatively on a fused SPECT/CT scan. In
19 the remainder of the manuscript these cases are referred to as 'CT' and 'SPECT/CT' respectively. All volumes of

1 interest (VOI) were exported to file in the DICOM format for further analysis. All VOIs were validated by a
2 physician.

3 **Clinical dosimetry reporting was done in accordance with EANM guidelines [21].**

4

5

6 Registration of sequential SPECT/CT images

7 In this study, we used both rigid and deformable image registration to bring, for each clinical case, the 5 sequential
8 SPECT/CT scans in the same frame of reference of the first CT scan.

9 The rigid registration was manually performed using the Siemens E-Soft workstation. Each SPECT scan was
10 registered to the reference CT scan using translations and rotations only, using an iterative process until the best match
11 for the kidneys was visually found.

12 The deformable registration was performed with Velocity (Varian Medical Systems, Palo Alto, USA), **which uses a**
13 **modified B-spline deformable algorithm with mutual information-based matching [22]. This algorithm performs a 3-**
14 **pass deformable registration (coarse, medium and fine resolution), that adds a very precise touch to the images. This**
15 **kind of deformation is recommended generally used for CT to CT registration and, in general, to register high**
16 **resolution images.** Velocity's deformable image registration algorithm was found to be capable of considerably
17 reducing the residual misalignment of anatomical structures by Hoffmann et al [23].

18 **In this work, the CT component of each SPECT/CT scan was first manually registered to the reference CT scan to**
19 **match the bony anatomy and then the Velocity deformable multi-pass registration algorithm was applied for optimal**

1 image fusion over the whole image. Images registered with the Velocity deformable registration algorithm were
2 visually assessed and considered to be clinically acceptable. The deformable registration was carried out between the
3 CT scans, and the resulting deformation map was applied to the corresponding SPECT scan. The temporal scan
4 sequence is reported in Figure 1 (a). Both rigid and deformable registration workflows are depicted in Figure 1 (b) and
5 Figure 1 (c). An example of a rigid and a deformable registration is shown in Figure 2 with the relevant VOIs.
6 The quality of the image registration was evaluated using the Structural Similarity Index (SSIM) [14] as implemented
7 in the MATLAB (The Mathworks, Natick MA) platform. SSIM was designed to provide an objective metric for
8 comparing a distorted image to a distortion-free (reference) image and is calculated as a combination of pixel
9 intensity, contrast and structural information [24]. In this work, SSIM was calculated in the range [0,100] where 100
10 indicates a perfect match between the images. For all the clinical cases, the SSIM values were calculated between the
11 first CT scan (used as reference) and the sequential CT scans registered using both rigid and deformable algorithms.
12 Differences between SSIM values for the deformable and rigid registration were compared with paired samples
13 Wilcoxon tests (two-sided, significance level 0,05) of the four comparison (images acquired at T2, T3, T4 and T5
14 compared to the CT acquired at T1). Statistical analysis was performed using R 3.3.3 (R Core Team. R: A Language
15 and Environment for Statistical Computing. R Foundation for Statistical Computing, Vienna, Austria, 2017).

16 17 Dosimetry calculations

18 For each patient, all registered images and VOIs were imported in Raydose [25], a full 3D Monte Carlo based
19 application for dosimetry calculations in MRT. Raydose is based on the general purpose Geant4 toolkit and can

1 transport all particles on a virtual phantom generated from a patient specific CT scan. Raydose can replicate the decay
2 of ^{177}Lu using data taken from the evaluated nuclear structure database (ENSDF available at:
3 www.nndc.bnl.gov/ensdf), and simulates beta-decay spectra and correct intensity photon emission in the subsequent
4 daughter nuclei which contributes to the overall absorbed dose [26]. Each SPECT scan, representing a voxel-by-voxel
5 snapshot of activity at a given time point, was used by Raydose to calculate a total absorbed dose map following the
6 method described by Marcatili et al [25]. 3D absorbed dose calculations using Raydose were carried out for both
7 rigidly and deformably registered scans (cf. Figure 1 (b) and (c) respectively) on two 20 core Dell PowerEdge rack
8 server (Intel Xeon E5-2670 @ 2.5GHz) with each machine equipped with 64 Gb RAM. Full patient absorbed dose
9 calculation, including 5 scans per patient, was carried out in approximately 9 hours. For each scan, 150 million events
10 were simulated giving a MC uncertainty < 5% in the total absorbed dose distribution map. Other sources of
11 uncertainty related for instance to activity calibration were not relevant to this work and were not included. The total
12 absorbed dose distribution calculated with Raydose was exported to file in the DICOM RTDOSE format.

14 Data Analysis

15 Standard radiotherapy tools such as Dose Volume Histograms (DVH) were used to assess the quality of the absorbed
16 dose distribution in all clinical cases and for both rigid and deformable image registration. Furthermore, minimum,
17 maximum and average absorbed dose to OARs and lesions were also calculated. A dosimetric comparison between
18 the two registration methods was carried out using the percentage difference (P.D.) figure of merit, calculated using
19 mean absorbed dose values in the VOIs:

1

2 $P.D.=(D_{rig}-D_{def})/D_{def} \times 100$

Eq 1

3

4 where D_{rig} and D_{def} are the mean **absorbed doses** calculated on rigidly registered images and deformably registered
5 images respectively. **The deformable image registration method was arbitrarily chosen as a reference given the**
6 **longitudinal nature of the study.** A positive P.D. indicates an overestimation of the average **absorbed dose** as
7 calculated using rigid registration, while a negative P.D. indicates an underestimation of **absorbed dose** from rigid
8 registration compared to **absorbed dose** from deformable image registration.

9

10 **Results**

11 **Figure 3 shows the SSIM values for the deformable registration (straight line) and for the rigid registration (dotted**
12 **line). All the differences between SSIM values were highly statistically significant ($p < 0.001$).**

13 **Single values for SSIM are reported in Table 1 for each patient and for each CT image (acquired at T2, T3, T4 and**
14 **T5), in comparison with the CT acquired at T1. Standard deviations for rigid and deformable registrations were also**
15 **calculated. It can be noted that the quality of the registration is higher in the case of deformably registered volumes.**

16 **Table 2 reports the **absorbed dose** volume statistics for kidneys normalized to the administered activity, and the**
17 **percentage difference, as defined in Eq 1. It can be noted that P.D. is negative for 10 out of 11 patients and **ranges****
18 **from +4% down to -19% (mean=-5,1%; median=-5,2%; 25th percentile=-7,6; 5th percentile=-8,9%). This indicates that**
19 **the average **absorbed dose** calculated using rigid image registration (D_{rig}) is consistently lower than the average**

1 absorbed dose calculated using the deformable algorithm (D_{def}). It is also worth noting that in some cases (e.g. patient
2 3 and 10) there is a large difference ($>5\%$) in P.D. between left and right kidney for the same patient, mainly due to a
3 difficult rigid registration, caused by patient's twisting or tilting.

4 Table 3 reports the absorbed dose volume statistics for lesions, normalized to the administered activity, and the
5 percentage differences. In this case P.D. ranges from -67.2% to 100.7% (mean= -15.0% ; median= -17.11% ; 25th
6 percentile= -6.25% ; 5th percentile= $+0.10\%$). The image modality used for the segmentation of the lesion is also
7 reported.

8 The DVHs calculated for rigid and deformable registration are depicted in Figure 4a (kidneys) and Figure 4b (lesion)
9 for clinical case No. 5 (cf. Figure 2).

10 Figure 5 depicts the P.D. as a function of volume for all VOIs. The imaging modality used to contour the lesions is
11 also reported.

12

13 **Discussion**

14 The analysis of the similarity index showed that the deformed CT scans are more similar to the reference CT than the
15 rigidly registered CT scans. The deformation matrices, or the rigid registration matrices are the same applied to NM
16 scans, so similarities in CT images are translated to similarities in NM scans. CT images are only taken as reference
17 for both registration algorithms, since NM scans are not adequate showing a time variant distribution of activity in
18 sequential NM imaging.

1 Hence, the deformable registration algorithm has a higher degree of improvement than the rigid registration, providing
2 an image more similar to the reference one. Furthermore, the standard deviations reported in Table 1 show also that
3 the deformable registration images produce a SSIM index much more stable than the rigid registration. This may be
4 expected, since the rigid registration may be more or less complex depending on the patient repositioning for
5 acquisition or patient movement during acquisitions. While rigid registration is intrinsically ‘locally’ accurate,
6 deformable image registration is ‘globally’ accurate as it succeeds in accounting for multiple internal organ movement
7 simultaneously.

8 The mean absorbed dose calculated with deformable image registration (D_{def}) was found to be in general higher than
9 the mean absorbed dose calculated with rigid registration (D_{rig}). This was expected, since the deformable registration
10 of images causes a change in the quantification of the activity inside VOIs after realignment. The negative value of
11 P.D. in most cases show that deformable registration provides a gain in activity quantification in the VOIs with a
12 consequently higher absorbed dose estimation. In our investigation, in 10 out of 11 cases, D_{rig} for kidneys was lower
13 than D_{def} by up to 19% as shown in Table 2. In 20 out of 22 cases, D_{rig} for lesions was lower than D_{def} by up to 67% as
14 shown in Table 3. An interesting correlation between P.D. and imaging modality in which the lesion was outlined
15 (‘CT’ or ‘SPECT/CT’) is also shown in Table 3. Indeed, it can be noted that when the lesion was outlined on a fused
16 scan (i.e. ‘SPECT/CT’) and the corresponding mass used in absorbed dose calculations, the P.D. for D_{rig} compared to
17 D_{def} was higher than 20%, in 80% of the cases. On the other hand, when the contour was based on CT and the
18 corresponding mass used in absorbed dose calculations, the P.D. was lower than 20%, in 75% of cases. This suggests
19 that the choice of the imaging modality used for target delineation is an additional important factor to consider when

1 the aim is to evaluate the effect of the deformable algorithm on the **absorbed dose** calculation. These results could be
2 helpful in informing volume segmentation protocols in future clinical trials in PRRT and other kind of MRT.
3 Figure 4 clearly shows that while the effect of the image registration algorithm on the **absorbed dose** statistics
4 extracted from relatively large OARs might be negligible, it can have a large impact on the **absorbed dose** calculation
5 for target lesions with underestimations of the **absorbed dose** to 50% of the volume that can be as large as 60% and an
6 underestimation of the maximum **absorbed dose** to the lesion up to 70%. This can have important consequences in
7 **absorbed dose** prescription and in the evaluation of the effectiveness of therapy.

8 Figure 5 shows a clear dependence of P.D. with lesion volume: larger volumes correspond to smaller P.D. In our
9 study, lesion volumes were in the range [4, 166] mL with a mean value of 34 mL. The same dependence is not so
10 evident for the kidneys (range [125, 346] mL, mean 214 mL). For larger lesions with a volume comparable to that of
11 the kidneys, the P.D. was lower than 5%, which is in line with the mean P.D. calculated in the analysis of the kidney
12 **absorbed dose** distribution. Our data indicates that deformable registration affects organ **absorbed dose** calculation
13 mainly when volumes are smaller than 100mL. This volume interval is in our experience the range of volume of
14 lesions in PRRT.

15 The following additional observations can be made: (a) in patient No. 5 (cf. Figure 2), 'liver lesion 1' (volume 8 mL,
16 'CT' contouring modality) is characterized by a large negative P.D. of -67.2% (figure 2 to see the area of the liver).
17 This value can be explained by the presence of a motion artefact in the abdomen hugely corrected using deformable
18 image registration, and by the modality of contouring; (b) in patient No. 9, 'liver lesion 3' (volume 26 mL, 'CT'
19 contouring modality) is characterized by a large positive P.D. value of 100.7%. This was due to a significant motion

1 artefact affecting the superior abdominal region causing a large mismatch between the reference CT image and the
2 sequential SPECT images. The deformable registration algorithm was unable to correct for the mismatch; (c) in
3 patient No. 6, 'liver lesion 4' (volume 4 mL, 'SPECT/CT' contouring modality) is characterized by a small P.D. of
4 0.3; in this case, the performance of both registration algorithms was equally good due to the reproducibility of the
5 location of the lesion across scans together with the limited spatial resolution of SPECT images.

6 In our experience, deformable image registration was a useful tool in accounting for misregistrations due to
7 respiratory motion and patient repositioning. The extent and the effect of the correction depended on many factors
8 including the location of the anatomical volume that needed to be registered, the anatomical position of the organ or
9 lesion and on the volume of the VOIs. Whereas several works about the deformable registration in external radiation
10 therapy have been published with plenty of results (e.g. in the work of Pukala J et al [27]), in MRT only few papers
11 are available [2, 13]. Also, a recent work shows the superiority of deformable registration of PET/CT to planning CT
12 by multiple positional metrics, compared to rigid registration [28].

13 Jackson et al. [2] reported that the accuracy of the CT to CT deformable image registration of SPECT/CT studies and
14 the limited spatial resolution of the SPECT camera may have a low impact on the estimated absorbed dose calculation
15 of large organs and of lesions down to 2-3cm in diameter (corresponding to spherical volumes between 4 and 14 mL).
16 This is true, because the exact voxel-to-voxel alignment have a marginal impact on estimated absorbed dose. In our
17 study, all lesions had a volume larger than 4mL and 16 out of 22 had a volume larger than or equal to 14 mL. In our
18 experience, deformable registration could be successfully used across the range of VOIs considered in this study. Even
19 if the differences between D_{rig} and D_{def} were found to be smaller for kidneys than for lesions, deformable image

1 registration proved to have a key role in the 3D absorbed dose calculation process as it took into account variations in
2 patient specific anatomical characteristics.

3 The differences between D_{rig} and D_{def} observed in our study are on average smaller than those observed by Ao et al
4 [13] in their study. This may be due to the fact that Ao et al used one single CT image acquired at 24h p.i. for the
5 attenuation correction in the reconstruction of each sequential SPECT image acquired at different time points. In
6 addition, they outlined VOIs on the SPECT scans, that in many cases are larger than VOIs contoured on CT images.

7 On the other hand, we acquired a different CT scan for each SPECT scan and used this information in both attenuation
8 correction and image registration. This approach makes quantitative imaging more accurate and more adequate for a
9 PRRT trial. Our image registration workflow was based on anatomical imaging and VOIs were outlined in the
10 majority of cases on the first CT scan.

11 In particular, in this study we also considered the contribution of ‘CT’ contouring modality in SPECT quantification
12 on big down to small clinical volumes (i.e. organ down to lesion size).

13 The ‘CT’ contouring modality may be a more accurate criterion to define the morphological size of tumours. Also,
14 Uribe et al [29] studied the CT guided segmentation of SPECT images in phantoms, compared to 40% fixed threshold
15 segmentation and to a complex segmentation algorithm developed by Grimes et al. [30]. They concluded that the ‘CT’
16 contouring modality performed better than SPECT 40% fixed threshold segmentation for objects with volume larger
17 than about 8.5ml, but depending on the shape of the inserts. For smaller volumes, they stated that the ‘CT’ contouring
18 modality in SPECT quantification is strongly affected by partial volume effect. The strong need for partial volume

1 effect correction (by calculation of recovery coefficient for missing activity) for small volumes obtained with a ‘CT’
2 contouring was reported also by Grassi et al [19].

3 Unfortunately, the ‘CT’ contouring modality may be difficult to use in clinical SPECT exams sometimes, since
4 tumours are not always clearly visible in CT images. A way to deal with this problem may be to adopt a contouring
5 based on SPECT images, which performs similarly to ‘CT’ contouring modality for small volumes [29]. This was
6 observed in this work as well.

7 In conclusion, three-dimensional image based dosimetry for ^{177}Lu -DOTATOC peptide receptor radionuclide therapy
8 is significantly affected by the type of algorithm used to register sequential SPECT/CT scans. We have shown that it is
9 feasible to implement in the clinical practice a workflow based on deformable image registration. This could have
10 important implications in the design of future trials in PRRT.

11 **Acknowledgments**

12 NP and ES acknowledge funding from the European Metrology Research Programme joint research project
13 “Metrology for Molecular Radiotherapy”, (MetroMRT) which has received funding from the European Union on the
14 basis of Decision No. 912/2009/EC.

17 **Conflicts of interest**

18 The authors have no relevant conflicts of interest to disclose.

1 References

2 [1] Ilan E, Sandström M, Wassberg C et al. *Dose response of pancreatic neuroendocrine tumors treated with*
3 *peptide receptor radionuclide therapy using ¹⁷⁷Lu-DOTATATE*. J Nucl Med. 2015 Feb;56(2):177-82.

4 [2] Jackson PA, Beauregard JM, Hofman MS et al. *An automated voxelized dosimetry tool for radionuclide*
5 *therapy based on serial quantitative SPECT/CT imaging*. Med Phys. 2013 Nov;40(11):112503.

6 [3] Kletting P, Schimmel S, Haenscheid H, et al. *The NUKDOS software for treatment planning in molecular*
7 *radiotherapy*. Z. Med. Phys. 2015 25: 264-274

8 [4] Baechler S, Hobbs RF, Boubaker A, Buchegger F, He B, Frey EC, Sgouros G.

9 *Three-dimensional radiobiological dosimetry of kidneys for treatment planning in peptide receptor radionuclide*
10 *therapy*. Med Phys. 2012 Oct;39(10):6118-28

11 [5] Woliner-van der Weg W, Schoffelen R, Hobbs RF, Gotthardt M, Goldenberg DM, Sharkey RM, Slump CH,
12 van der Graaf WT, Oyen WJ, Boerman OC, Sgouros G, Visser EP.

13 *Tumor and red bone marrow dosimetry: comparison of methods for prospective treatment planning in pretargeted*
14 *radioimmunotherapy*. EJNMMI Phys. 2015 Dec;2(1):5

15 [6] D'Arienzo M, Capogni M, Smyth V, et al. *Metrological Issues in Molecular Radiotherapy* EPJ Web of
16 Conferences 77, 00022 (2014) <http://dx.doi.org/10.1051/epjconf/20147700022>

17 [7] Gustafsson J, Brolin G, Cox M et al. *Uncertainty propagation for SPECT/CT-based renal dosimetry in*
18 *(¹⁷⁷Lu) peptide receptor radionuclide therapy*. Phys Med Biol. 2015 Nov 7;60(21):8329-46.

- 1 [8] Guerra L, De Ponti E, Elisei F et al. *Respiratory gated PET/CT in a European multicentre retrospective*
2 *study: added diagnostic value in detection and characterization of lung lesions.* Eur J Nucl Med Mol Imaging. 2012
3 Sep; 39(9):1381-90.
- 4 [9] Verhey LJ *Immobilizing and Positioning Patients for Radiotherapy* Seminars in Radiation Oncology, Vol 5,
5 No 2 (April), 1995:pp 100-114
- 6 [10] Kutcher GJ, Coia L, Gillin M et al. *Comprehensive QA for radiation oncology: report of AAPM radiation*
7 *therapy committee task group 40,* Med.Phys. 21 (4) , 1994
- 8 [11] Papavasileiou P, Flux GD, Guy MJ et al. *A novel four-dimensional image registration method for*
9 *radionuclide therapy dosimetry.* Phys Med Biol. 2004 Dec 7;49(23):5373-91.
- 10 [12] Papavasileiou P, Divoli A, Hatzioannou K et al. *A generalized 4D image registration scheme for targeted*
11 *radionuclide therapy dosimetry.* Cancer Biother Radiopharm. 2007 Feb;22(1):160-5.
- 12 [13] Ao EC, Wu NY, Wang SJ et al. *Improved dosimetry for targeted radionuclide therapy using*
13 *nonrigid registration on sequential SPECT images.* Med Phys. 2015 Feb;42(2):1060-70.
- 14 [14] Brock KK, Mutic S, McNutt TR et al. *Use of image registration and fusion algorithms and techniques in*
15 *radiotherapy: Report of the AAPM Radiation Therapy Committee Task Group No. 132.* Med Phys. 2017 July, 44(7):
16 e43-e76.
- 17 [15] Bodei L, Mueller-Brand J, Baum RP et al. *The joint IAEA, EANM, and SNMMI practical guidance on peptide*
18 *receptor radionuclide therapy (PRRNT) in neuroendocrine tumours.* Eur J Nucl Med Mol Imaging. 2013
19 May;40(5):800-16.

- 1 [16] Sghedoni R, Grassi E, Fioroni F et al. *Personnel exposure in labelling and administration of (177)Lu-DOTA-*
2 *D-Phe1-Tyr3-octreotide*. Nucl Med Commun. 2009, Feb;30(2):176-82.
- 3 [17] Uribe CF, Esquinas PL, Gonzalez M, Celler A *Characteristics of Bremsstrahlung emissions of*
4 *177Lu, 188Re, and 90Y for SPECT/CT quantification in radionuclide therapy*. Phys. Med. 2016;32(5):691-
5 700
- 6 [18] Grassi E, Fioroni F, Ferri V, et al. *Quantitative comparison between the commercial software STRATOS® by*
7 *Philips and a homemade software for voxel-dosimetry in radiopeptide therapy*. Phys Med. 2015 Feb;31(1):72-9.
- 8 [19] Grassi E, Fioroni F, Mezzenga E, et al. *Impact of a commercial 3D OSEM reconstruction algorithm on the*
9 *177Lu activity quantification of SPECT/CT imaging in a Molecular Radiotherapy trial*. Radiol. Diagn. Imaging 2017;
10 1(1): 1-7.
- 11 [20] Bailey, Dale L, Kathy P Willowson. "Quantitative SPECT/CT: SPECT joins PET as a quantitative imaging
12 modality." *Eur J Nucl Med Mol Imaging* (Springer) 41, no. 1 (May 2014): 17-25.
- 13 [21] Lassmann M, Chiesa C, Flux G, et al. *EANM dosimetry committee guidance document: good practice of*
14 *clinical dosimetry reporting*. Eur J Nucl Med Mol Imaging 2011;38:192-200.
- 15 [22] Lawson JD, Schreibmann E, Jani AB et al. *Quantitative evaluation of a cone-beam computed tomography-*
16 *planning computed tomography deformable image registration method for adaptive radiation therap*. J Appl Clin
17 *Med Phys*. 2007 Nov 5;8(4):2432.
- 18 [23] Hoffmann C, Krause , Stoiber E M et al. *Accuracy quantification of a deformable image registration tool*
19 *applied in a clinical setting*. J Appl Clin Med Phys. 2014 Jan 6;15(1):4564.

- 1 [24] Zhou, W., A. C. Bovik, H. R. Sheikh, and E. P. Simoncelli. "Image Quality Assessment: From Error Visibility
2 to Structural Similarity." *IEEE Transactions on Image Processing*. Vol. 13, Issue 4, April 2004, pp. 600–612
- 3 [25] Marcatili, S, C Pettinato, S Daniels et al. "*Development and validation of RAYDOSE: a Geant4-based*
4 *application for molecular radiotherapy.*" *Phys. Med. Biol.* 58 (2013): 2491.
- 5 [26] Hauf S, Kuster M, Batič M et al. *Radioactive Decays in Geant4*, *IEEE Trans. Nuc. Sci.* 2013, 60(4). 2966
- 6 [27] Pukala J, Johnson PB, Shah AP et al *Benchmarking of five commercial deformable image registration*
7 *algorithms for head and neck patients.* *J Appl Clin Med Phys.* 2016 May;17(3):25-40. doi: 10.1120/jacmp.v17i3.5735.
- 8 [28] Ward G, Ramasamy S, Sykes JR et al *Superiority of Deformable Image Co-registration in the Integration of*
9 *Diagnostic Positron Emission Tomography-Computed Tomography to the Radiotherapy Treatment Planning Pathway*
10 *for Oesophageal Carcinoma.* *Clin Oncol (R Coll Radiol).* 2016 Oct;28(10):655-62.
- 11 [29] Uribe CF, Esquinas PL, Tanguay J et al Accuracy of ¹⁷⁷Lu activity quantification in SPECT imaging: a
12 phantom study *EJNMMI Physics* 2017 4:2 Published on: 7 January 2017
- 13 [30] Grimes J, Celler A, Shcherbinin S, et al The accuracy and reproducibility of SPECT target volumes and
14 activities estimated using an iterative adaptive thresholding technique. *Nucl Med Commun.* 2012;33:1254–66.

1 **Table 1**

2 **Brief Title for Table 1: SSIM values**

Case Number	CT(T2) versus CT(T1)		CT(T3) versus CT(T1)		CT(T4) versus CT(T1)		CT(T5) versus CT(T1)	
	Deformable	Rigid	Deformable	Rigid	Deformable	Rigid	Deformable	Rigid
1	99,58%	98,53%	99,59%	97,18%	99,44%	96,44%	99,62%	96,44%
2	99,66%	98,95%	99,62%	98,16%	99,60%	98,47%	99,62%	98,17%
3	99,62%	97,86%	99,38%	98,30%	99,56%	98,01%	99,30%	97,98%
4	99,62%	98,49%	99,66%	96,99%	99,55%	97,71%	99,62%	98,54%
5	99,61%	98,60%	99,59%	98,76%	99,71%	98,63%	99,37%	97,81%
6	99,53%	98,71%	99,40%	98,27%	99,48%	97,14%	99,38%	97,97%
7	99,66%	97,35%	99,55%	97,08%	99,43%	97,11%	99,55%	95,32%
8	99,73%	97,04%	99,65%	97,09%	99,69%	97,97%	99,67%	98,25%
9	99,65%	97,32%	99,65%	97,92%	99,62%	98,61%	99,55%	98,03%
10	99,57%	98,95%	99,49%	98,34%	99,40%	98,12%	99,42%	97,61%
11	99,49%	97,67%	99,36%	96,98%	99,37%	96,53%	99,44%	96,88%
Standard deviation	0,07%	0,70%	0,11%	0,67%	0,12%	0,79%	0,13%	0,96%

3 **Table 1** SSIM values calculated for each patient and for each CT image (acquired at T2, T3, T4 and T5) in
4 comparison with the CT acquired at T1. Index values were used to investigate the similarity of deformably and rigidly
5 registered images with the reference CT acquired at T1.

6
7
8
9
10

1 **Table 2**

2

3

4 **Brief Title for Table 2:** Absorbed dose volume statistics for kidneys

5

6

7

8

9

10

11

12

13

14

15

Case number	VOI		Absorbed dose (mGy/MBq)			Absorbed dose (mGy/MBq)			P.D.
			Deformable registration			Rigid registration			
			Name	Vol (mL)	Min	Mean	Max	Min	
1	right kidney	268	0.03	0.51	1.25	0.04	0.49	1.18	-4.7
	left kidney	210	0.01	0.50	1.30	0.03	0.49	1.52	-0.4
2	right kidney	135	0.13	0.82	3.21	0.06	0.76	3.23	-7.9
	left kidney	144	0.00	0.65	1.29	0.00	0.59	1.24	-8.9
3	right kidney	138	0.10	0.73	1.40	0.07	0.67	1.40	-8.1
	left kidney	125	0.07	0.65	1.58	0.05	0.64	1.55	-2.6
4	right kidney	283	0.01	0.37	1.14	0.02	0.37	1.13	-0.5
	left kidney	282	0.02	0.39	1.21	0.01	0.37	1.14	-5.6
5	right kidney	156	0.22	1.08	2.04	0.20	1.05	1.86	-2.6
	left kidney	156	0.10	1.01	2.07	0.09	0.94	2.04	-6.9
6	right kidney	346	0.01	0.73	1.80	0.01	0.72	1.75	-1.3
	left kidney	252	0.07	0.84	1.98	0.05	0.80	1.86	-4.6
7	right kidney	198	0.06	0.76	1.71	0.05	0.71	1.71	-6.1
	left kidney	196	0.04	0.69	1.69	0.04	0.65	1.89	-6.4
8	right kidney	241	0.09	0.77	1.58	0.07	0.71	1.37	-7.9
	left kidney	254	0.03	0.74	1.43	0.02	0.67	1.38	-8.8
9	right kidney	203	0.01	0.84	1.93	0.02	0.87	1.77	4.2
	left kidney	174	0.06	0.90	2.10	0.06	0.89	2.01	-0.4
10	right kidney	234	0.03	0.47	0.95	0.02	0.38	0.99	-19.0
	left kidney	232	0.02	0.47	0.98	0.03	0.45	1.32	-4.0
11	right kidney	243	0.05	0.58	1.20	0.05	0.55	1.28	-5.1
	left kidney	246	0.02	0.56	1.24	0.02	0.53	1.24	-5.3

- 1 **Table 2 Absorbed dose** volume statistics for kidneys normalized to the administered activity and percentage
- 2 difference on mean **Absorbed dose** values (cf. Eq 1) between rigid and deformable registration.

3

1 **Table 3**

2 **Brief Title for Table 3:** Absorbed dose volume statistics for lesions

3

4

5

6

7

8

9

10

11

12

13

14

15

16

17

Case	Lesion		Absorbed dose (mGy/MBq)			Absorbed dose (mGy/MBq)			P.D.	Modality	
	Type	No.	Vol (mL)	Deformable registration			Rigid registration				
				Min	Mean	Max	Min	Mean			Max
1	liver	1	166	0.18	2.52	5.21	0.05	2.20	5.05	-12.6	CT
	liver	2	148	0.09	1.60	4.04	0.10	1.53	3.71	-4.7	CT
2	liver	1	23	0.21	2.79	8.03	0.37	2.36	7.52	-15.6	CT
	liver	2	20	1.06	4.46	6.83	0.66	4.15	6.72	-7.0	CT
3	liver	1	10	0.24	0.78	1.71	0.20	0.61	1.69	-21.5	SPECT/CT
4	bone	1	39	0.00	0.18	0.85	0.01	0.14	0.61	-21.6	SPECT/CT
5	liver	1	8	0.22	0.76	1.74	0.07	0.25	0.59	-67.2	CT
6	lung	1	20	0.05	2.37	7.26	0.11	1.93	6.79	-18.6	CT
	liver	2	15	0.30	1.64	4.44	0.25	1.29	3.92	-21.5	SPECT/CT
	liver	3	25	0.18	1.45	3.41	0.05	0.96	3.45	-33.5	SPECT/CT
	liver	4	4	0.19	0.61	1.37	0.05	0.61	1.40	0.3	SPECT/CT
7	abdomen	1	6	0.16	0.36	0.50	0.14	0.34	0.50	-6.1	CT
	abdomen	2	38	0.10	0.52	1.19	0.12	0.49	1.09	-6.1	CT
8	liver	1	11	0.13	0.86	2.72	0.07	0.48	2.66	-44.2	SPECT/CT
	liver	2	6	0.12	0.58	1.35	0.07	0.32	1.13	-43.9	SPECT/CT
	liver	3	14	0.12	0.73	2.86	0.10	0.70	2.55	-3.8	SPECT/CT
9	abdomen	1	44	0.08	2.26	10.07	0.13	2.11	7.77	-6.7	CT
	abdomen	2	16	0.25	5.29	16.73	0.34	4.77	16.30	-9.9	CT
	liver	3	26	0.19	1.50	9.54	0.20	3.01	10.95	100.7	CT
10	abdomen	1	56	0.04	0.98	2.64	0.07	0.71	2.81	-27.0	CT
	liver	2	47	0.09	0.77	1.59	0.14	0.57	1.66	-26.4	SPECT/CT
	liver	3	26	0.15	0.76	2.82	0.12	0.51	1.96	-32.8	SPECT/CT

1

2

3 **Table 3 Absorbed dose** volume statistics for lesions normalized to the administered activity and percentage

4 **difference (cf. Eq 1) between rigid and deformable registration. The imaging modality used for VOI**

5 **segmentation is also reported.**

6

7

8

1 **Captions**

2

3 **Figure 1** Temporal sequence (a), rigid registration (b) and deformable registration (c) of SPECT/CT scans acquired in
4 this study. Identity registrations are represented by the double line connector (=). Rigid and deformable registrations
5 are represented by the straight () and curved line connectors (*) respectively. All transformation matrices operating
6 on a given scan, can also be applied to all other scans that are in the same frame of reference (i.e. identity registered).

7

8 **Figure 2** Example of rigid (a, b, c, d) and deformable (e, f, g, h) registration for clinical case No. 5. Kidneys are
9 visible in the transverse axis (a, e) while the lesion is shown in the transverse (b, f), coronal (c, g) and sagittal axis (d,
10 h).

11

12 **Figure 3** SSIM value calculated for deformable registrations (solid line) and rigid registrations (dotted line). CT
13 images acquired at T2, T3, T4 and T5 were separately compared with the CT acquired at T1 ($p < 0.001$ for each plot).

14

15 **Figure 4** DVHs for kidneys (a) and lesions (b) calculated on rigidly (dashed line) and deformably (solid line)
16 registered SPECT/CT scans for clinical case No. 5 (cf. Figure 2).

17

- 1 **Figure 5** Percentage Difference on mean organ absorbed dose as a function of volume for all clinical cases considered
- 2 in this study.